

FINAL VERSION

VERSION FINALE



**Medical electrical equipment –
Part 2-33: Particular requirements for the basic safety and essential performance
of magnetic resonance equipment for medical diagnosis**

**Appareils électromédicaux –
Partie 2-33: Exigences particulières pour la sécurité de base et les performances
essentielle des appareils à résonance magnétique utilisés pour le diagnostic
médical**

CONTENTS

FOREWORD.....	4
* INTRODUCTION	7
INTRODUCTION TO AMENDMENT 1	8
INTRODUCTION TO AMENDMENT 2	8
201.1 Scope, object and related standards.....	9
201.2 Normative references	10
201.3 Terms and definitions	11
201.4 General requirements.....	18
201.5 General requirements for testing of ME EQUIPMENT.....	18
201.6 Classification of ME EQUIPMENT and ME SYSTEMS	18
201.7 ME EQUIPMENT identification, marking and documents.....	18
201.8 Protection against electrical HAZARDS from ME EQUIPMENT	31
201.9 Protection against mechanical HAZARDS of ME EQUIPMENT and ME SYSTEMS	32
201.10 Protection against unwanted and excessive radiation HAZARDS.....	33
201.11 Protection against excessive temperatures and other HAZARDS.....	33
201.12 Accuracy of controls and instruments and protection against hazardous outputs	33
201.13 HAZARDOUS SITUATIONS and fault conditions	53
201.14 PROGRAMMABLE ELECTRICAL MEDICAL SYSTEM (PEMS)	53
201.15 Construction of ME EQUIPMENT	53
201.16 ME SYSTEMS.....	53
* 201.17 Electromagnetic compatibility of ME EQUIPMENT and ME SYSTEMS	54
* 202 Electromagnetic compatibility – Requirements and tests.....	54
Annexes	56
Annex D (informative) Symbols on marking.....	57
Annex AA (informative) Particular guidance and rationale	61
Bibliography.....	114
Index of defined terms used in this particular standard.....	125
Figure 201.101 – Gradient waveform and EFFECTIVE STIMULUS DURATION	12
Figure 201.102 – Limits for cardiac and peripheral nerve stimulation	37
Figure 201.103 – Reduction of WHOLE BODY SAR limits at high temperatures	41
Figure 201.104 – Volume for determining the spatial maximum of gradient output	47
Figure 201.105 – Volume for determining the B_1 stray field	50
Figure AA.1 – Static magnetic fields: flow potentials and retardation.....	84
Figure AA.2 – Experimental data on PNS threshold of human volunteers in WHOLE BODY MR EQUIPMENT.....	99
Figure AA.3 – Double logarithmic plot of experimental threshold values for peripheral nerve stimulation	100
Figure AA.4 – Response value $R(t)$ generated by convolution of a rectangular stimulus dB/dt and a nerve impulse response function $n(t-\theta)$	104
Figure AA.5 – Gradient waveform G , stimulus waveform dB/dt and response value R , for a trapezoid EPI waveform starting at $t = 0$	105

Figure AA.6 – Threshold values dB/dt for two gradient waveforms, plotted against EFFECTIVE STIMULUS DURATION	105
Figure AA.7 – Threshold value of dB/dt for a sinusoid gradient waveform, as function of the number of half periods in the waveform.....	106
Figure AA.8 – SAR limits for the exposed mass of a PATIENT	109
Table 201.101 – List of symbols and abbreviations	17
Table 201.102 – Rheobase values per type of gradient system	37
Table 201.103 – Weight factors for summation of the maximum output O_j per GRADIENT UNIT	38
Table 201.104 – Temperature limits	38
Table 201.105 – SAR limits for volume transmit coils	39
Table 201.106 – SAR limits for local transmit coils	40
Table 201.107 – FPO limits applicable for cylindrical MR SYSTEMS	51
Table 201.D.101 – MR SAFETY SIGNS	57
Table 201.D.102 – RF coil symbols.....	59
Table 201.D.103 – MR conditional symbols.....	60
Table AA.1 – Overview physiological effects in humans, animals and model systems, for magnetic-field exposures at field strengths relevant for MRI. The implications of the reported effects are assessed in the last column.....	75

INTERNATIONAL ELECTROTECHNICAL COMMISSION

MEDICAL ELECTRICAL EQUIPMENT –**Part 2-33: Particular requirements for the basic safety and essential performance of magnetic resonance equipment for medical diagnosis**

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

DISCLAIMER

This Consolidated version is not an official IEC Standard and has been prepared for user convenience. Only the current versions of the standard and its amendment(s) are to be considered the official documents.

This Consolidated version of IEC 60601-2-33 bears the edition number 3.2. It consists of the third edition (2010-03) [documents 62B/777/FDIS and 62B/782/RVD], its amendment 1 (2013-04) [documents 62B/884/CDV and 62B/904/RVC] and its amendment 2 (2015-06) [documents 62B/977/FDIS and 62B/987/RVD]. The technical content is identical to the base edition and its amendments.

This Final version does not show where the technical content is modified by amendments 1 and 2. A separate Redline version with all changes highlighted is available in this publication.

International standard IEC 60601-2-33 has been prepared by IEC subcommittee 62B: Diagnostic imaging equipment, of IEC technical committee 62: Electrical equipment in medical practice.

This edition of IEC 60601-2-33 is based on the second amendment to Edition 2. It has also been adapted to the third edition of IEC 60601-1 (2005), with technical modifications being introduced where appropriate.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

In this standard, the following print types are used:

- Requirements and definitions: roman type.
- *Test specifications: italic type.*
- Informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type.
- TERMS DEFINED IN CLAUSE 3 OF THE GENERAL STANDARD, IN THIS PARTICULAR STANDARD OR AS NOTED: SMALL CAPITALS.

In referring to the structure of this standard, the term

- “clause” means one of the seventeen numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 7 includes subclauses 7.1, 7.2, etc.);
- “subclause” means a numbered subdivision of a clause (e.g. 7.1, 7.2 and 7.2.1 are all subclauses of Clause 7).

References to clauses within this standard are preceded by the term “Clause” followed by the clause number. References to subclauses within this particular standard are by number only.

In this standard, the conjunctive “or” is used as an “inclusive or” so a statement is true if any combination of the conditions is true.

The verbal forms used in this standard conform to usage described in Annex H of the ISO/IEC Directives, Part 2. For the purposes of this standard, the auxiliary verb:

- “shall” means that compliance with a requirement or a test is mandatory for compliance with this standard;
- “should” means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this standard;
- “may” is used to describe a permissible way to achieve compliance with a requirement or test.

An asterisk (*) as the first character of a title or at the beginning of a paragraph or table title indicates that there is guidance or rationale related to that item in Annex AA.

A list of all parts of the IEC 60601 series, published under the general title: *Medical electrical equipment*, can be found on the IEC website.

The committee has decided that the contents of the base publication and its amendments will remain unchanged until the stability date indicated on the IEC web site under "<http://webstore.iec.ch>" in the data related to the specific publication. At this date, the publication will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

The contents of the corrigendum of March 2012 have been included in this copy.

IMPORTANT – The 'colour inside' logo on the cover page of this publication indicates that it contains colours which are considered to be useful for the correct understanding of its contents. Users should therefore print this document using a colour printer.

* INTRODUCTION

This particular standard is written at a moment in which the technical evolution of MR EQUIPMENT is in rapid progress and the scientific foundation of its safe use is still expanding.

This International Standard addresses technical aspects of the medical diagnostic MR SYSTEM and the MR EQUIPMENT therein related to the safety of PATIENTS examined with this system, the safety of the MR WORKER involved with its operation and the safety of the MR WORKER involved with the development, manufacturing, installation, and servicing of the MR SYSTEM. Where limits of electromagnetic fields (EMF) exposure of PATIENTS and MR WORKERS are stated, these limits do not imply that such levels of exposure can be assumed to be acceptable for workers in other professional settings and for the population at large. The limits provide a sensible balance between RISKS for the PATIENTS and MR WORKERS and benefits for the PATIENTS.

Organizational aspects of safety are the task of the RESPONSIBLE ORGANIZATION. This task includes adequate training of staff, rules of access to the MR SYSTEM, qualification of staff for decisions that are related to safety, definition of medical responsibility and specific requirements for personnel following from that responsibility when the PATIENT is in or near the MR SYSTEM.

Examples of such organizational aspects are:

- operation in FIRST LEVEL CONTROLLED OPERATING MODE;
- emergency procedures for resuscitation of the PATIENT who is in the MR SYSTEM;
- emergency procedures after a QUENCH of the superconductive magnet when present;
- set-up and maintenance of a protocol for screening the PATIENT for contraindications or for conditions that may affect acceptable exposure;
- rules for ROUTINE MONITORING and for MEDICAL SUPERVISION of the PATIENT during the exam.
- rules to minimize and to limit the exposure of MR WORKERS to EMF.

Extensive rationale is provided in Annex AA for some of the definitions and requirements in order to provide the user of this standard with a reasonably complete access to the source material that was used in support of the considerations during drafting.

The relationship of this particular standard with IEC 60601-1 and the collateral standards is explained in subclauses 201.1.3 and 201.1.4.

The introduced EMF exposure limits required in this standard for an MR WORKER will never exceed those allowed for PATIENTS. All exposure limits allowed for a PATIENT and for an MR WORKER are expected to protect them against negative health effects and unacceptable RISKS.

For the exposure to static magnetic fields, subjective short-term physiological and sensory effects are expected. These influence the well being of the MR WORKER marginally and only during or shortly after exposure.

For the exposure to GRADIENT OUTPUT and RF transmit fields, normally no short-term physiological and sensory effects are expected for MR WORKERS.

In addition no experimental or theoretical basis for cumulative biological effects in humans, resulting from exposure at the allowed levels has been generally accepted.

The requirements for acoustic noise exposure are different for PATIENTS and MR WORKERS.

INTRODUCTION TO AMENDMENT 1

This amendment has been published to adapt IEC 60601-2-33:2010 to the technical corrections introduced by Amendment 1 (2012) to IEC 60601-1:2005.

INTRODUCTION TO AMENDMENT 2

This Amendment 2 has been developed to increase the FIRST LEVEL CONTROLLED OPERATING MODE limit for the static field from 4 T to 8 T taking into account FDA, ICNIRP and other peer reviewed scientific literature. In addition, a non-compulsory option, FIXED PARAMETER OPTION:BASIC (FPO:B), is introduced to limit RF and gradient field outputs (peak and RMS) for scanning PATIENTS with MR conditional implants. Consequently, text is proposed for the Instructions for use to guide users in scanning PATIENTS with MR conditional implants.

Furthermore, references to newly published collateral standards have been updated.

MEDICAL ELECTRICAL EQUIPMENT –

Part 2-33: Particular requirements for the basic safety and essential performance of magnetic resonance equipment for medical diagnosis

201.1 Scope, object and related standards

Clause 1 of the general standard¹⁾ applies, except as follows:

201.1.1 Scope

Replacement:

This International Standard applies to the BASIC SAFETY and ESSENTIAL PERFORMANCE of MR EQUIPMENT and MR SYSTEMS, hereafter referred to also as ME EQUIPMENT.

This standard does not cover the application of MR EQUIPMENT beyond the INTENDED USE.

If a clause or subclause is specifically intended to be applicable to ME EQUIPMENT only, or to ME SYSTEMS only, the title and content of that clause or subclause will say so. If that is not the case, the clause or subclause applies both to ME EQUIPMENT and to ME SYSTEMS, as relevant.

The standard does not formulate specific requirements for MR EQUIPMENT or MR SYSTEMS used in INTERVENTIONAL MR EXAMINATIONS.

201.1.2 Object

Replacement:

The object of this particular standard is to establish particular BASIC SAFETY and ESSENTIAL PERFORMANCE requirements for MR EQUIPMENT to provide protection for the PATIENT and the MR WORKER.

NOTE This standard presumes that the MR WORKERS are properly medically screened, and properly trained and instructed in their duties.

201.1.3 Collateral standards

Addition:

This particular standard refers to those applicable collateral standards that are listed in Clause 2 of the general standard and Clause 201.2 of this particular standard.

IEC 60601-1-2:2014 applies as modified in Clause 202. IEC 60601-1-3, IEC 60601-1-10, IEC 60601-1-11 and IEC 60601-1-12 do not apply. All other published collateral standards in the IEC 60601-1 series apply as published.

201.1.4 Particular standards

Replacement:

¹⁾ The general standard is IEC 60601-1:2005+A1:2012, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*